



CONGRATULATIONS ON YOUR PURCHASE OF A PRODUCT FROM NOUVAG.

We are pleased that you have chosen a quality product from NOUVAG and thank you very much for the trust you have placed in us.

These instructions for use will familiarize you with the device and its functions so that you can apply and use them correctly.

SYMBOLS



General warning sign



General mandatory action



Refer to instructions for use



Manufacturer



Date of manufacture



Importer



Do not use if the package is damaged



Not for reuse



Use-by date



Biological hazard



Not made with natural rubber latex



Contains or presence of Phthalate



Batch code



Catalog number



Serial number



Medical device



Sterilized using ethylene oxide



Authorized representative in the European Community



Water resistance



Equipotentiality




Indication of pump flow direction



Type BF applied part



Foot switch



European Conformity mark



Separate collection required (WEEE)



Certified by the TÜV Rheinland North America Group

CONTENT

PRODUCT DESCRIPTION	4
Intended use and operation	
Contraindications	
Ambient conditions	
SAFETY INFORMATION	5
EMC Manufacturer's Declaration of Conformity	
Possible risks and side effects	
Modifications and misuse	
Essential requirements	
During use	
SCOPE OF DELIVERY	7
DEVICE OVERVIEW	9
Front view	
Rear view	
SETUP	10
Device setup	
Connection to the power supply	
Potential equalization connection according to DIN 42801	
Preparing reusable secret jars	
Preparing secretion jars with disposable inlay pouch	
Device preparation	
OPERATION	12
Switching the device on and off	
Activating the pump via the pneumatic pedal	
Varying with VARIO-AIR pedal	
Regulating the suction process	
Emptying secretion jars	
Function check	
CLEANING AND DISINFECTION	15
Control unit and pneumatic pedal	
Secretion jar and jar lid	
Secretion jar with disposable inlay pouch	
Bacteria filter	
Cannulas and cannula handle	
Quiver and silicone tubes	
MAINTENANCE	17
Replacing the control unit fuses	
Safety inspections	
Bacteria filter	
Secretion jar	
Function control of float gauge valve	
MALFUNCTIONS AND TROUBLESHOOTING	18
ACCESSORIES AND SPARE PARTS	19
Information on disposal	
TECHNICAL DATA	20
SERVICE	21
Post market surveillance	
Service points	
APPENDIX	22

PRODUCT DESCRIPTION

INTENDED USE AND OPERATION

The Vacuson serves as a suction pump and is used in the following field of applications:

- ↪ Surgery pump
- ↪ Lipectomy pump or subcutaneous liposuction
- ↪ Universal pump

The Vacuson function is to aspirate fluids and secretions. The pump's suction power can be adjusted continuously using a vacuum controller and monitored using the manometer.

Patient population is not restricted in respect of age, weight and gender.

Configuration and operation of the Vacuson shall be performed only by surgeons or highly qualified and trained medical personnel.

CONTRAINDICATIONS

Infectious wounds Liposuction may only be performed after the treatment of the infection and necrotic tissue.

In principle, generally poor health of the patient.

Liposuction shortly after a strict diet of the patient.

Morbid obesity (obesity) Large suction volumes increase the risk of death due to fluid shifts.

Relevant cases in the literature must be considered.

AMBIENT CONDITIONS

	TRANSPORT AND STORAGE	DURING USE
Relative humidity	10%–90%	max. 80%
Temperature	0 °C–50 °C	10 °C–30 °C
Atmospheric pressure	700 hPa–1'060 hPa	800 hPa–1'060 hPa

SAFETY INFORMATION

It is essential to bear the following information in mind:

Every use of the Vacuson different to the product description defined in section [\[INTENDED USE AND OPERATION >4\]](#) causes risks for patients and trained personnel. If physical examinations and therapies are carried out without use of the devices then the device must be removed from the place of treatment.

EMC MANUFACTURER'S DECLARATION OF CONFORMITY

The use of (RF) Radio Frequency emitting devices and equipment as well as the occurrence of negative environmental factors in the close area of the Vacuson may cause unexpected or adverse operation. The connection or the placing of other devices in close vicinity is not allowed.

Use only accessories and cables as specified in the product description. Further observe the EMC manufacturer declaration of conformity.

POSSIBLE RISKS AND SIDE EFFECTS

- Improper use can result in tissue or organ injuries to the patient or cuts to the user or a third person.
- In rare cases, a treatment can lead to mild neurological disorders. In very rare cases, a treatment can lead to endovenous heat-induced thrombosis.

SAFETY INFORMATION

MODIFICATIONS AND MISUSE



Modifications/manipulations on the Vacuson and its accessories are not permitted. Failure to follow these instructions can have unpredictable consequences for the user, the patient or third parties. For consequential complications, resulting from illicit modifications/manipulations the manufacturer assumes no responsibility and the guarantee is void.

ESSENTIAL REQUIREMENTS



Do not use the device if the shipping box has holes/cracks on the flat surfaces, and/or if the Styrofoam protective packaging is broken.

The Vacuson may only be operated by qualified and trained personnel!

The use of third-party products is the responsibility of the operator. Functionality and patient safety cannot be guaranteed with third-party accessories.

Repairs may only be performed by authorized NOUVAG service technicians!

Improper use or repair of the device, or failure to observe these instructions, relieves NOUVAG from any obligation arising from warranty provisions or other claims.

Prior to using the device, before startup, and before operation, the user must always ensure that the device and accessories are in good working order and are clean, sterile and operational.

Ensure that the operating voltage setting corresponds to the local mains voltage.

DURING USE



The device is not sterile on delivery. Please observe the instructions [[CLEANING AND DISINFECTION >15](#)].

At choice of the instrument the user has to make sure it confirms to EN ISO 10993, means that it's biocompatible.

Do not use device in the vicinity of flammable mixtures!

While in operation the control unit of the Vacuson suction pumps must be at least 1 meter above ground.

In extreme cases, the device may heat up excessively.

The use of the Vacuson other than that for which it was designed (see [[INTENDED USE AND OPERATION >4](#)]) is not permitted. The responsibility is solely carried by the operator.

SCOPE OF DELIVERY

VACUSON 40 SET (REF 4227-115 / REF 4227-230)

REF	DESCRIPTION	QUANTITY
4275-115	Vacuson 40 (115V) Control unit	1
4275-230	Vacuson 40 (230V) Control unit	
15012	Foot pedal ON/OFF, pneumatic, IPX8	1
4076	Suction tube Ø8 × 3 × 1700 mm, silicone, sterilisable	1
4155z	Connecting tube Ø8 × 3 × 400 mm, from bacteria filter to secretion jar, silicone, sterilisable	1
4246	Bacteria filter for suction pump, Ø64 mm, PTFE, hydrophobe, disposable	10
31997	Vacuson 40 Vacuson 60 Instructions for use	1

VACUSON 60 SET (REF 4237-115 / REF 4237-230)

REF	DESCRIPTION	QUANTITY
4280-115	Vacuson 60 (115V) Control unit	1
4280-230	Vacuson 60 (230V) Control unit	
15012	Foot pedal ON/OFF, pneumatic, IPX8	1
4076	Suction tube Ø8 × 3 × 1700 mm, silicone, sterilisable	1
4155z	Connecting tube Ø8 × 3 × 400 mm, from bacteria filter to secretion jar, silicone, sterilisable	1
4246	Bacteria filter for suction pump, Ø64 mm, PTFE, hydrophobe, disposable	10
31997	Vacuson 40 Vacuson 60 Instructions for use	1

OPTIONAL

REF	DESCRIPTION	QUANTITY
4030F	FLOVAC® suction jar system, for disposable bags, 2 litres, with mounting bracket	1
4035F	Disposable inlay pouches, 2 litres, including lid, for the FLOVAC® system	50
4019F	FLOVAC® tubing adapter (yellow), for attachment on the VACUUM connection on the lid of the inlay pouch	25
4036F	Inlay pouches, 2 litres, for the FLOVAC® system, sterilisable	1
4037F	Mounting bracket for the FLOVAC® system, for mounting secretion jars on Vacuson 40/60 pump	1
4044	Quiver, sterilisable, 400 mm length, with suspension device	1
4130	Two way tap for switching between the secretion jars, including connection tube Ø8 × 3 × 400 mm	1
4190	Connecting tube Ø 8 × 3 × 500 mm, from bacteria filter to secretion jar, silicone, sterilisable	1
4242	VARIO-AIR Pedal	1
6026 6026E	Disposable tubing set Ø9 × 6.5 mm, sterile, 4 m, PU 40 pcs.	1
4052nou 4052usa	Secretion jar, 2 litre, polysulfone, sterilisable	1
4245nou 4245usa	Secretion jar, 5 litre, polysulfone, sterilisable	1
4287	Suction jar 2 l, reusable	1
4288	Suction jar 5 l, reusable	1
4289	Lid for suction jars 2 l / 5 l, reusable	1
4291	Suction jar 3 l, reusable	1
4292	Suction liner 3 l, single-use, PU 70 pcs.	1
4290	Coupling piece 90°, reusable, PU 10 pcs.	1

SCOPE OF DELIVERY

OPTIONAL POWER CORDS

REF	DESCRIPTION	QUANTITY
22261	Power cord CH, with device socket, 3 m	1
22262	Power cord DE, with device socket, 3 m	1
22264	Power cord GB, with device socket, 3 m	1
22266	Power cord US, with device socket, 3 m	1


DEVICE OVERVIEW

FRONT VIEW




REAR VIEW

VACUSON 40



VACUSON 60



1 Suction cannula (optional) 2 Suction tube (1700 mm), silicone 3 Secretion jar (example 2 litres) with lid and float system 4 Connection for connecting tube (VACUUM) 5 Connection for connecting tube (PATIENT) 6 Connecting tube, silicone 7 Bacteria filter 8 Two way cock (optional, REF 4130) 9 Carrying handle 10 Secretion jar mount 11 Ready indication, LED 12 Manometer 13 Pneumatic ON/OFF-pedal 14 Vacuum regulator «VACUUM» 15 Ventilation intake 16 Potential equalization 17 Port for pneumatic ON/OFF-pedal 18 Air outlet port «EXHAUST» 19 Connection for VARIO-Air pedal 20 Power plug socket 21 Main switch «ON/OFF» 22 Fuse compartment 23 Type plate with type designation, reference number, serial number, information on power supply and device fuse

SETUP

DEVICE SETUP

INSTALLATION LAYOUT



- Place the Vacuson suction pump and all required accessories and instruments on an even, non-slip surface and make sure you have good access to all controls.
- Do not allow the operating range of the device (including the cable) to be compromised by limiting factors.
- The manometer must be fully visible at all times.
- The foot switch must be placed within stepping distance between the patient and the surgeon.
- It must be explicitly ensured that no objects can fall onto the foot switch.
- The power plug at the rear of the device must be accessible at all times.
- The ventilation slots at the housings bottom and sideways of the Vacuson must be kept clear in order to prevent temperature from becoming excessive.
- While in operation the Vacuson suction pumps must be at least 1 meter above ground.

CONNECTION TO THE POWER SUPPLY



Before switching on, make sure that the power supply unit of the device matches the country's specific service voltage!

In order to prevent the risk of electric shock, the device may only be connected to a power network with a PE protective earth conductor.

Only a tested mains cable may be used to connect the device to the power supply.

POTENTIAL EQUALIZATION CONNECTION ACCORDING TO DIN 42801

At the back of the device a potential equalization plug is installed, according to DIN 42801.

The additional potential equalization has the task of equalizing potentials between different parts of conductive materials that can be touched at the same time, or reducing potential differences.

This connection must be used, to protect the patient, the user and third parties from touch voltages.

The equipotential plug is marked with the following symbol: 

SETUP

PREPARING REUSABLE SECRET JARS

- 1 Have an open, sterile secretion jar (2 or 5 litres) ready.



Follow the instructions for use when installing the secretion jars.

PREPARING SECRETION JARS WITH DISPOSABLE INLAY POUCH

- 1 Hold the secretion jar ready with disposable inlay pouch and mounting bracket.
- 2 Place the angle connector on the lid (PATIENT), place the insert bag in the jar and close the jar by applying sustained pressure to the lid.
- 3 Hang the holder ring on the Vacuson and attach the jar.
- 4 Plug the connecting tube via the tube adapter (REF 4019F) onto the VACUUM tube and the suction tube onto the PATIENT connection.

DEVICE PREPARATION



Prior to use all sterilizable parts (silicone tubes, cannulas, cannula handle and secretion jars) must be sterilised.

- 1 Hang the secretion jars with the lid on and closed in the secretion jar holder on the side of the device.
- 2 Place the ON/OFF pneumatic pedal on the floor and insert the connecting tube into the socket provided at the back of the device.
- 3 The VARIO-AIR-pedal (optional) is connected to the connection on the back of the device via the connecting tube (only for Vacuson 40).



When the VARIO-AIR pedal is not in use, be sure to replace the cover, otherwise no vacuum can be created.

- 4 Attach the short piece of silicone tubing (400 mm) to the secretion jar (VACUUM). Connect the other end of the tube to the bacteria filter.
- 5 Attach one end of the suction tube (1700 mm silicone tube) to the secretion jar (PATIENT) and attach the instrument to the other end.
- 6 Connect the suction cannula (optional) to the cannula handle and hang the cannula with handle in the quiver.
- 7 Connect the mains plug to the power supply.

OPERATION

SWITCHING THE DEVICE ON AND OFF

The control unit is switched on and off using the main switch «I/O» on the back of the device. Readiness is indicated by a green LED light on the front of the device.

Switching off can take place at any time and is not dependent on a switch-off procedure.



The ON/OFF pneumatic pedal must be connected, as the pump only runs when the pedal is pressed.

ACTIVATING THE PUMP VIA THE PNEUMATIC PEDAL

When pressed, the ON/OFF pneumatic pedal supplied with the Vacuson triggers a switch in the device to switch the pump on or off. This is done via an air cushion. After switching on the device, press the pneumatic pedal.

Pump runs and generates vacuum. Pressing the pedal again switches the pump off. The pump can only be activated with the pedal. The last state before switching off remains active.



It may not be possible to switch on the pump when a vacuum has been generated. In this case, release the vacuum via the vacuum regulator and press the pneumatic pedal again.

VARYING WITH VARIO-AIR PEDAL

If a VARIO-AIR pedal (REF 4242) was also ordered with your Vacuson device, it can be used to regulate the vacuum. When pressed, the VARIO-AIR pedal opens a valve that controls the supply of ambient air. The further the pedal is depressed, the more ambient air is drawn in and the weaker the suction power at the cannula. When using the VARIO-AIR pedal to control the suction, the vacuum regulator must be in the maximum position. The VARIO-AIR pedal is used together with the pneumatic pedal.

REGULATING THE SUCTION PROCESS

The suction process is controlled by the vacuum regulator on the front of the device.

Turning clockwise Vacuum increases, suction performance is enhanced

Turn anticlockwise Vacuum decreases, suction performance is reduced

SETTING THE DESIRED SUCTION POWER

- 1 **Turn the vacuum regulator to the left stop.**
Corresponds to minimum vacuum.
- 2 **Switch on the device using the main switch «I/O» and briefly press the ON/OFF pneumatic pedal.**
The pump is running and vacuum is being built up.
- 3 **Crimp suction tube to generate maximal air-tightness of the suction system.**
- 4 **Turn the vacuum control to the right stop.**
Corresponds to maximum vacuum.
- 5 **Wait until the pump has built up maximum vacuum.**
Equivalent to $\geq -0,9$ bar.
- 6 **By turning the vacuum control, the suction power can now be adjusted continuously from 0 to $-0,9$ bar.**



The manometer indicates the current vacuum in the device. Due to the connection of tubes and adapters, the effective vacuum at the cannula may differ from the displayed value.

OPERATION

EMPTYING SECRETION JARS

The lid of the secretion jar is equipped with an overflow protection. This prevents suction of secretion liquid when the jar is full. Therefore a float gauge is responsible. If the liquid level is too high and the overflow protection closes, the secretion jar must be emptied or replaced with an empty one.

- 1 Switch off the device using the main switch «I/O».
- 2 Disconnect the tubes from the lid of the secretion jar.
- 3 Unhinge the full secretion jar from the secretion jar mount and dispose of the secretion fluids according to national disposal regulations.
- 4 Procure the used secretion jars to the reprocessing cycle.
- 5 Connect the disconnected tubes to the second secretion jar that is ready and continue the suction procedure.

FUNCTION CHECK

To ensure that the suction pump operates smoothly, the individual components and functions must be checked before each treatment.

GENERAL FUNCTION

- 1 Switch on the device using the main switch «I/O».
Standby display (LED) is lit.
The fan in the device is running.

SUCTION PUMP WITH PNEUMATIC PEDAL

- 1 Switch on the device using the main switch «I/O».
Standby display (LED) is lit.
The fan in the device is running.
- 2 Hold hand in front of the air outlet «EXHAUST» (back of the device).
Airflow can be felt.
- 3 Activate the suction pump with the pneumatic pedal provided.
- 4 Turn the vacuum controller «Vacuum» to the right stop to get the maximum suction power.
- 5 Check the suction intensity at the cannula opening.
The pump draws strongly.
- 6 Turn the vacuum regulator «Vacuum» to the left stop to get the minimum suction power.
- 7 Check the suction intensity at the cannula opening.
The pump draws weakly.
- 8 Switch off the device via the main switch «I/O».
Standby display (LED) goes out.

OPERATION

SUCTION PUMP WITH PNEUMATIC PEDAL AND VARIO-AIR-PEDAL

- 1 Switch on the device using the main switch «I/O».**
Standby display (LED) is lit.
The fan in the device is running.
- 2 Hold hand in front of the air outlet «EXHAUST» (back of the device).**
Airflow can be felt.
- 3 Activate the suction pump with the pneumatic pedal provided.**
- 4 Turn the vacuum controller «Vacuum» to the right stop to get the maximum suction power.**
- 5 Check the suction intensity at the cannula opening.**
The pump draws strongly.
- 6 Press the VARIO-AIR-Pedal.**
The more the pedal is pressed, the weaker the suction intensity at the cannula becomes.
- 7 Turn the vacuum regulator «Vacuum» to the left stop to get the minimum suction power.**
- 8 Check the suction intensity at the cannula opening.**
The pump draws weakly.
- 9 Press the VARIO-AIR-Pedal.**
The more the pedal is pressed, the weaker the suction intensity at the cannula becomes.
With the vacuum regulator «Vacuum» in the minimum position, the suction intensity can no longer be felt.
- 10 Switch off the device via the main switch «I/O».**
Standby display (LED) goes out.

CLEANING AND DISINFECTION



Clean and disinfect after every treatment!

Always autoclave the material in sterilisation packaging.

Sterilisation packaging may only be filled up to 80%.

Always autoclave the material at 134°C for at least 5 minutes.

If sterilised material is not used immediately, the material packaging must be labeled with the sterilisation date.

NOUVAG recommends including a sterility indicator.

CONTROL UNIT AND PNEUMATIC PEDAL

Control unit and pneumatic pedal do not come into contact with the patient. Wipe the outside using tested surface disinfectant or 70% of isopropyl alcohol. The cover of the device is appropriately sealed and can be wiped clean.

SECRETION JAR AND JAR LID

The reprocessing instructions for the secretion jar and jar lid is provided in the operating instructions delivered with the secretion jar.

SECRETION JAR WITH DISPOSABLE INLAY POUCH

The disposable inlay pouches are not to be reprocessed. They have to be discarded off expertly. For reprocessing the reusable outer container (secretion jar) please refer to the operation manual delivered with the product.

BACTERIA FILTER

The bacteria filter is a one way product and cannot be cleaned or sterilised. A periodical replacement of the bacteria filter is recommended after 8 hours of use, but definitely after it came into contact with foam or infectious material.



After contact with watery solutions the bacteria filter locks down, because of its hydrophobic characteristic, to protect the pump from clogging. Hence the further operation of the pump is not possible. The bacteria filter has to be replaced.

CANNULAS AND CANNULA HANDLE

The optional cannulas and the cannula handle are in contact with the patient and therefore have to be reprocessed adequately. The reprocessing instructions are in the operation instructions, delivered together with the cannula and handle.

QUIVER AND SILICONE TUBES

Clean quiver from debris and soiling. Use a clean, damp cloth and/or an appropriate brush with disinfection agent. Flush tubes sufficiently.



It is important to use a disinfectant that is compatible with polysulfone and polycarbonate.

Note: Quiver and silicone tubes must be disposed of after 250 reprocessings at the latest.

- 1 Pack quiver and tubes in individual packaging for sterile items (see DIN 58953).
- 2 Autoclave wrapped quiver and tubing at 134°C for at least 5 minutes.
Temperature exposure times are based on country-specific guidelines and standards.
- 3 A drying cycle must be added in case of autoclaves without a post-vacuum function. Allow quiver to dry in the bag for at least one hour at room temperature with the paper side facing upwards.



If sterilised quiver and tubing are not used immediately after sterilisation, the material packaging must be labeled with the sterilisation date. Including a sterility indicator is recommended.

CLEANING AND DISINFECTION

Reprocessing restrictions	Frequent reprocessing has little effect on the products. The end of product life is usually determined by wear and tear and damage through use.
----------------------------------	---

INSTRUCTIONS

At location of use	No special requirements.
Storage and transport	No special requirements. Long holding times before reprocessing have to be avoided due to surface drying.
Preparation for cleaning	No special requirements.
Automatic cleaning and disinfection	Equipment Washer-disinfector with a special load carrier that ensures the connection of tubes to the washer-disinfector for rinsing. Use only neutral cleaning agents for this purpose. 1. Place silicone tubes in the load carrier. 2. Set a cleaning cycle that offers sufficient cleaning and rinsing. Perform the final rinse with fully deionized water. 3. Perform a 10-minutes rinse cycle at 93°C to facilitate thermal disinfection. 4. When removing, check silicone tubes, to verify whether soiling is still visible. If necessary, repeat the cycle or clean manually.
Manual cleaning	Equipment Neutral cleaning agent, soft brush, running, demineralized water (< 38°C) Procedure 1. Rinse off and brush away surface soiling from the silicone tubes. 2. Rinse silicone tubes thoroughly under running water.
Manual disinfection	For manual disinfection, submerge silicone tubes in chlorinefree disinfection solution.
Drying	Allow silicone tubes to dry sufficiently in a drying cabinet.
Inspection and maintenance	Perform a visual inspection to check for damage, corrosion and wear.
Packaging	Individual Pack silicone tubes in individual packaging for sterile items. Sets Sort silicone tubes on trays intended for this purpose or place them on allpurpose sterilisation trays.
Sterilisation	Autoclave in vacuum autoclave at 134°C for at least 5 minutes. When sterilizing several items during one sterilisation cycle, do not exceed the maximum sterilizer load. A drying cycle must be added in case of autoclaves without a post-vacuum function. Allow the silicone tubes to dry in the bag for at least one hour at room temperature with the paper side facing upwards. * Temperature exposure times are based on country-specific guidelines and standards.
Storage	No special requirements. If sterilized silicone tubes are not used immediately after sterilisation, the material packaging must be labeled with the sterilisation date. Including a sterility indicator is recommended.


The effectiveness of the sterilisation instructions provided above for reprocessing this medical product has been validated by NOUVAG AG. The user is responsible for ensuring that the sterilisation procedure performed achieves the required results. This requires validation and routine monitoring of the procedure. The staff member who completes the procedure bears sole responsibility for any deviation on his part from the instructions provided. Deviations necessitate revalidation of the effectiveness of the procedure as well as of the technical resilience of the reprocessed items with regard to the modified sterilisation process.

MAINTENANCE

REPLACING THE CONTROL UNIT FUSES

Users can replace faulty control unit fuses themselves. These are located at the rear of the device in the fuse slot next to the power switch:

- 1 Switch off device.
- 2 Unplug the power cord.
- 3 Open the fuse slot using a screwdriver.
- 4 Replace the faulty fuse T 4AL, 250V AC (115V model) / T 2AL, 250V AC (230V model).
- 5 Slide the fuse holder back in and close the fuse slot.
- 6 Plug in the power cord again.



1 Fuse slot locking mechanism 2 Display window for voltage setting 3 Fuse slot 4 Fuse 1 5 Fuse 2

SAFETY INSPECTIONS

The essential requirements have been defined and assessed within the risk analysis. The results of the analysis are stored in the risk management file of the manufacturer.

The performance of safety inspections on medical devices is required by law in several countries. The safety inspection is a regular safety check that is compulsory for those operating medical devices. The objective is to ensure that device defects and risks to patients, users or third parties are identified in time.

The STI (Safety Technical Inspection) for the Vacuson shall be executed every 2 years by authorized experts. Results shall be documented. The service manual, wiring diagrams, and descriptions are available upon request from the manufacturer.

NOUVAG offers a safety inspection service for its customers. Addresses can be found in the appendix of this instructions for use under [\[SERVICE POINTS >21\]](#). For further information please contact our technical service department.

BACTERIA FILTER

A periodical replacement of the bacteria filter is recommended after 8 hours of use, but definitely after it came into contact with foam or infectious material. For reordering refer to chapter [\[ACCESSORIES AND SPARE PARTS >19\]](#).

SECRETION JAR

The influxing mixture of air and secretion fluids into the secretion jar causes the build up of foam. It's recommended to use an antifoam agent to suppress the build up of foam. Prior use of the secretion jar fill an anti foam agent into the clean, dry jar. Don't use disinfection solution, because most of them benefit the build up of foam. Make sure the secretion jars are in good condition. Check the jars routinely for cracks and rifts and be sure the jars flange is immaculate. It's important to guarantee full air tightness of the system which is responsible for trouble-free operation of the pump.

FUNCTION CONTROL OF FLOAT GAUGE VALVE

The proper functioning of the overflow protection system has to be checked periodically.



Please refer to the instructions for use of the secretion jars.

MALFUNCTIONS AND TROUBLESHOOTING

MALFUNCTION	CAUSE	SOLUTION	REFER TO INSTRUCTIONS FOR USE
Device is not operational	Control unit not switched on	Set the power switch «I/O» to «I»	[SWITCHING THE DEVICE ON AND OFF >12]
	Power connection not established	Connect the control unit to the mains power supply	[CONNECTION TO THE POWER SUPPLY >10]
	Incorrect operating voltage	Check the mains voltage	[CONNECTION TO THE POWER SUPPLY >10]
	Faulty fuse	Replace fuse	[REPLACING THE CONTROL UNIT FUSES >17]
Pedal is not functioning	Pneumatic pedal not connected	Connect pneumatic pedal with device at rear	[ACTIVATING THE PUMP VIA THE PNEUMATIC PEDAL >12]
	Control unit is not switched on	Set the power switch «I/O» to «I»	[SWITCHING THE DEVICE ON AND OFF >12]
	Incorrect operation	Check instructions for use	
Suction pump does not work	Vacuum pump is not switched on	Press the connected pneumatic pedal	[ACTIVATING THE PUMP VIA THE PNEUMATIC PEDAL >12]
	Vacuum-system is not air tight	Check all seals and tubes Make sure char lid is properly closed	[PREPARING REUSABLE SECRET JARS >11] [PREPARING SECRETION JARS WITH DISPOSABLE INLAY POUCH >11] [DEVICE PREPARATION >11]
	Connection cap for connecting the VARIO-AIR pedal is not fitted	Place cap on pedal connection	[DEVICE PREPARATION >11]
	Tubes are connected wrong	Connect tubes correctly	[DEVICE PREPARATION >11]
	Jar is full and overflow protection has locked down	Replace full jar by a fresh, empty jar	[EMPTYING SECRETION JARS >13]
	Incorrect operation	Check instructions for use	
Suction pump does not suck properly	Vacuum controller «Vacuum» is not open enough	Turn vacuum controller «Vacuum» clockwise	[REGULATING THE SUCTION PROCESS >12]
	Vacuum-system is not air tight	Check all seals and tubes Make sure char lid is properly closed	[PREPARING REUSABLE SECRET JARS >11] [PREPARING SECRETION JARS WITH DISPOSABLE INLAY POUCH >11] [DEVICE PREPARATION >11]

If the problem cannot be solved please contact your supplier or an authorized service center. Addresses can be found in the appendix of this instructions for use under [SERVICE POINTS >21].

ACCESSORIES AND SPARE PARTS

DESCRIPTION	REF
O-ring Ø11.1 × 1.6 mm, silicone	4063
Connecting tube Ø8 × 3 × 400 mm, from bacteria filter to secretion jar, silicone, sterilisable	4155z
Suction tube Ø8 × 3 × 1700 mm, silicone, sterilisable	4076
Connecting tube Ø8 × 3 × 500 mm, from bacteria filter to secretion jar, silicone, sterilisable	4190
FLOVAC® suction jar system, for disposable bags, 2 litres, with mounting bracket	4030F
Disposable inlay pouches, 2 litres, including lid, for the FLOVAC® system	4035F
FLOVAC® tubing adapter (yellow), for attachment on the VACUUM connection on the lid of the inlay pouch	4019F
Inlay pouches, 2 litres, for the FLOVAC® system, sterilisable	4036F
Mounting bracket for the FLOVAC® system, for mounting secretion jars on Vacuson 40/60 pump	4037F
Bacteria filter for suction pump, Ø64 mm, PTFE, hydrophobe, disposable	4246
Two way tap for switching between the secretion jars, including connection tube Ø8 × 3 × 400 mm	4130
Disposable tubing set Ø9 × 6.5 mm, sterile, 4 m, PU 40 pcs.	6026/6026E
Secretion jar, 2 litre, polysulfone, sterilisable	4052nou / 4052usa
Secretion jar, 5 litre, polysulfone, sterilisable	4245nou / 4245usa
Suction jar 2 l, reusable	4287
Suction jar 5 l, reusable	4288
Lid for suction jars 2 l / 5 l, reusable	4289
Suction jar 3 l, reusable	4291
Suction liner 3 l, single-use, PU 70 pcs.	4292

SUCTION CANNULAS FOR LIPOSUCTION

DESCRIPTION	REF
Yankauer suction cannula, Ø2.0 mm, length 280 mm	4446
Andrews cannula, Ø2.0 mm, length 240 mm	4449
Cannula handle with opening for false air ventilation, Luer-Lock connection, sterilisable	4391
Cannula handle without opening, Luer-Lock connection, sterilisable	4390
Curved cannula, for femoral liposuction, Ø3 mm, length 200 mm, 22 openings 1.5 mm	4362
Curved cannula, for femoral liposuction, Ø3 mm, length 300 mm, 30 openings 1.5 mm	4365
Curved cannula, for femoral liposuction, Ø4 mm, length 200 mm, 22 openings 1.5 mm	4368
Curved cannula, for femoral liposuction, Ø4 mm, length 300 mm, 30 openings 1.5 mm	4372
Angled cannula, 30°, for femoral liposuction, Ø3 mm, length 200 mm, 22 openings 1.5 mm	4381
Straight cannula, Ø1.5 mm, length 150 mm, 1 oval opening	4361
Straight cannula, Ø2 mm, length 150 mm, 1 oval opening	4364
Straight cannula, Ø2 mm, length 150 mm, 18 opening 1.0 mm	4373
Straight cannula, Ø3 mm, length 150 mm, 18 openings 1.5 mm	4374
Straight cannula, Ø3 mm, length 200 mm, 22 openings 1.5 mm	4378
Straight cannula, Ø3 mm, length 300 mm, 30 openings 1.5 mm	4387
Straight cannula, Ø4 mm, length 200 mm, 22 openings 2.0 mm	4379
Straight cannula, Ø4 mm, length 300 mm, 30 openings 1.5 mm	4388

To order any additional parts, please contact our customer service department.

ACCESSORIES AND SPARE PARTS

INFORMATION ON DISPOSAL

When disposing of the device, device components and accessories, the regulations issued by the legislator must be followed.



Contaminated single-use tubing sets are subject to specific disposal requirements. Please observe prevailing national disposal regulations.

Disused electrical and electronic appliances are hazardous waste and must not be disposed of with household waste. Old appliances can be returned to the dealer or manufacturer.

TECHNICAL DATA

VACUSON 40

Voltage	115 V~ / 60 Hz, 230 V~ / 50 Hz	
Power consumption	115 V Model	max. 180 VA
	230 V Model	max. 170 VA
Fuse power supply	115 V Model	2 fuses, T 4AL, 250 V AC
	230 V Model	2 fuses, T 2AL, 250 V AC
Protection class	Class I	
Applied part	Type BF	
Adjustable vacuum	– 0.9 bar at 686 mmHg	
Suction capacity vacuum pump	35 l/min.	
Accuracy of pressure gauge	± 5 %	
Dimensions (W x D x H)	360 x 300 x 280 mm	
Net weight control unit	10 kg	

VACUSON 60

Voltage	115 V~ / 60 Hz, 230 V~ / 50 Hz	
Power consumption	115 V Model	max. 370 VA
	230 V Model	max. 400 VA
Fuse power supply	115 V Model	2 fuses, T 4AL, 250 V AC
	230 V Model	2 fuses, T 2AL, 250 V AC
Protection class	Class I	
Applied part	Type BF	
Adjustable vacuum	– 0.9 bar at 675 mmHg	
Suction capacity vacuum pump	60 l/min.	
Accuracy of pressure gauge	± 5 %	
Dimensions (W x D x H)	360 x 300 x 280 mm	
Net weight control unit	12 kg	

SERVICE

POST MARKET SURVEILLANCE



If you have any complaints in relation to the use of the medical device, please contact the manufacturer immediately by e-mail complaint@nouvag.com or by phone.

In order to provide adequate information, please complete the complaint form:

<https://nouvag.com/en/service/complaint-form/>

SERVICE POINTS



Switzerland

NOUVAG AG
St. Gallerstrasse 25
9403 Goldach

Phone +41 71 846 66 00
info@nouvag.com
nouvag.com



Germany

NOUVAG GmbH
Schulthaissstrasse 15
78462 Konstanz

Phone +49 7531 1290-0
info-de@nouvag.com
nouvag.com

A complete list of all NOUVAG authorised service points worldwide can be found at
<https://nouvag.com/en/service/global-service-centres/>

APPENDIX

Electromagnetic compatibility (EMC)

Remark:

The **Product** subsequently referred to herein always denotes the Vacuson 40 / 60.

Changes or modifications to this product not expressly approved by the manufacturer may result in increased emissions or decreased immunity performance of the product and could cause EMC issues with this or other equipment. This product is designed and tested to comply with applicable regulations regarding EMC and shall be installed and put into service according to the EMC information stated as follows.

WARNING

Use of portable phones or other radio frequency (RF) emitting equipment, including accessories (antennas e.g.) in distances below 30 cm (12 inches) to the product, may cause unexpected or adverse operation.

WARNING

The product is suitable for use in hospitals other than in the vicinity of active devices of the HF surgical devices or except in HF screening rooms used for magnetic resonance imaging.

WARNING

The product shall not be used adjacent to, or stacked with, other equipment. If adjacent or stacked use is necessary, the product shall be tested to verify normal operation in the configuration in which it is being used.

Essential Performance

The essential performance is the aspiration of aqueous secretions taking account of the suction rate set. The maximum suction rate deviation is $\pm 15\%$, the aspiration flowrate is between 5 and 55 l/min and the maximum vacuum is ≥ -0.95 bar.

Compliant Cables and Accessories

WARNING

The use of accessories, transducers and cables other than those specified may result in increased emissions or decreased immunity performance of the product.

Guidance and manufacturer's declaration – electromagnetic emissions		
The Product is intended for use in the electromagnetic environment specified below. The customer or the user of the Product should assure that it is used in such an environment.		
Emissions test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The Product uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. The Product is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
RF emissions CISPR 11	Class B	
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/flicker emissions IEC 61000-3-3	complies	

Guidance and manufacturer's declaration – electromagnetic immunity			
The Product is intended for use in the electromagnetic environment specified below. The customer or the user of the Product should assure that it is used in such an environment.			
Immunity tests	IEC 60601 Test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	+/- 6 kV contact +/- 8 kV air	+/- 6 kV contact +/- 8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	+/- 2 kV for power supply lines +/- 1 kV for input/output lines	+/- 2 kV for power supply lines +/- 1 kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	+/- 1 kV differential mode +/- 2 kV common mode	+/- 1 kV differential mode +/- 2 kV common mode	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	< 5 % U_T (> 95 % dip in U_T) for 0,5 cycle 40 % U_T (60 % dip in U_T) for 5 cycles 70 % U_T (30 % dip in U_T) for 25 cycles < 5 % U_T (> 95 % dip in U_T) for 5 sec	< 5 % U_T (> 95 % dip in U_T) for 0,5 cycle 40 % U_T (60 % dip in U_T) for 5 cycles 70 % U_T (30 % dip in U_T) for 25 cycles < 5 % U_T (> 95 % dip in U_T) for 5 sec	Mains power quality should be that of a typical commercial or hospital environment. If the user of the Product requires continued operation during power mains interruptions, it is recommended that the Product be powered from an uninterruptible power supply or a battery.
Power frequency (50/60Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
Note: U_T is the a.c. mains voltage prior to application of the test level.			

APPENDIX

Guidance and manufacturer's declaration – electromagnetic immunity for not life support equipment			
The Product is intended for use in the electromagnetic environment specified below. The customer or the user of the Product should assure that it is used in such an environment.			
Immunity tests	IEC 60601 Test level	Compliance level	Electromagnetic environment - guidance
			Portable and mobile RF communications equipment should be used no closer to any part of the Product, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance:
Conducted RF IEC 61000-4-6	3 V rms 150 kHz to 80 MHz outside ISM bands	3 V rms 150 kHz to 80 MHz outside ISM bands	$d = 0,35 \sqrt{P}$
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m 80 MHz to 2.5 GHz	$d = 0,35 \sqrt{P}$ 80 MHz to 800 MHz $d = 0,7 \sqrt{P}$ 800 MHz to 2,5 GHz
			where P is the maximum output power rating in the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey ^a , should be less than the compliance level in each frequency range ^b . Interference may occur in the vicinity of equipment marked with the following symbol: 
Note 1: At 80 MHz and 800 MHz, the higher frequency range applies.			
Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			
a	Fixed strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To access the electromagnetic environment due to fixed RF transmitters, and electromagnetic site survey should be considered. If the measured field strength in the location in which the Product is used exceeds the applicable RF compliance level above, the Product should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the Product.		
b	over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.		

Recommended separation distances between portable and mobile RF communications equipment and the not life support equipment			
The Product is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Product can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Product as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter m		
	150 kHz to 80 MHz $d = 0,35 \sqrt{P}$	80 MHz to 800 MHz $d = 0,35 \sqrt{P}$	800 MHz to 2.5 GHz $d = 0,7 \sqrt{P}$
0,01	0,04	0,04	0,07
0,1	0,11	0,11	0,22
1	0,35	0,35	0,7
10	1,1	1,1	2,2
100	3,5	3,5	7
For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the higher frequency range applies.			
Note 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.			
Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			



NOUVAG AG
St. Gallerstrasse 25
9403 Goldach
Switzerland

Phone +41 71 846 66 00
info@nouvag.com
nouvag.com



NOUVAG GmbH
Schulthaisstrasse 15
78462 Konstanz
Germany

Phone +49 7531 1290-0
info-de@nouvag.com
nouvag.com